

	<b>NATIONAL OIL CORPORATION OF KENYA</b> <b>ADDENDUM NO. 1 - SUPPLY OF COMPLETE SET OF FUEL GAS SAMPLING EQUIPMENT - 19TH APRIL 2024</b>	<b>Ref No. NOCK/PRC/03(1721)2023/24</b>
		<b>Page 1 of 1</b>

In reference to the **TENDER FOR SUPPLY OF COMPLETE SET OF FUEL GAS SAMPLING EQUIPMENT – REF NO. NOCK/PRC/03(1721)2023/24** published on 12th April 2024. We hereby issue addendum No.1 providing clarification as follows:

No.	REQUEST FOR CLARIFICATION	NATIONAL OIL RESPONSE TO THE REQUEST
1.	<p>Could you please provide clarity on the following points:</p> <p>Tank Size: We need confirmation on the preferred size of the tanks to be included in the equipment set. Could you specify whether the tanks should be measured in liters or gallons? Additionally, if there are specific size requirements or recommendations, kindly outline them for our reference.</p>	<p>National Oil provide clarification as follows:</p> <ol style="list-style-type: none"> <li>1. Gas Bags: Volume – 0.3 litre (300ml)</li> <li>2. Isotube or equivalent – approx. 110ml</li> <li>3. Gas Flask Volume – 750ml stand up pouch design</li> </ol>
	<p>Quantity of Tanks: It would be immensely helpful if you could specify the total number of tanks required as part of the complete set of Fuel Gas Sampling Equipment. Understanding the exact quantity will assist us in accurately preparing our proposal.</p>	<p>Ref to Price Schedule Form</p>

**Except as clarified herein, all the other terms and conditions of the tender remain as is and are unchanged.**

**CHIEF EXECUTIVE OFFICER**